



ESSCO CALIBRATION LABORATORY

DIVISION OF WALSH ENGINEERING SERVICES, INC.

CALIBRATION/REPAIR OF ALL TYPES OF ELECTRONIC / MECHANICAL TEST EQUIPMENT

27 INDUSTRIAL AVE UNIT #9, CHELMSFORD, MA 01824-3618

TEL: (800) 325-2201 (978) 250-0880 www.esscolab.com

CERTIFICATE OF CALIBRATION

ISSUE DATE: 06/10/2022

Certificate #: 2569408

Page 1 of 1

CUSTOMER / LOCATION APPLI-TEC 7 INDUSTRIAL WAY UNIT 1 SALEM, NH 03079 PURCHASE ORDER: 025281-00	EQUIPMENT INFORMATION MANUFACTURER: GENERIC MODEL NO: ESD FLOOR ONSITE SERIAL NO: N/A CONTROL NO: 0792 TYPE: ESD FLOOR
--	--

AS FOUND	IN TOLERANCE
AS LEFT	IN TOLERANCE

Remarks:
ON SITE

PERFORMED: ONSITE ENVIRONMENT: TEMPERATURE (deg C): 21
 METHOD: ECP NO. 1.4.42 Rev. 4 12/1/2020 REL HUMIDITY (%RH): 50

CALIBRATION DATE: 06/10/2022
 CALIBRATION DUE: 06/10/2023
METROLOGIST: 
 Jeff Lier

Standards Used To Calibrate Equipment

I.D.	Description	Report No.	Last Cal. Date	Cal. Due Date
E3193	DESCO 19787 SURFACE RESISTANCE METER KIT	2349801	7/22/2021	7/22/2022

The ESSCO Quality System is accredited to ISO/IEC 17025:2017

The results above relate only to the item(s) calibrated. This certificate shall not be reproduced, except in full, without written approval of ESSCO. The signature to the right signifies responsibility for the quality system.

This document certifies that the unit conformed to applicable specifications upon successful completion of the calibration. Any number of factors may cause the calibration item to drift out of calibration before the recommended interval has expired. The standards used are traceable to NIST or a National Measurement Institute.

The calibration was performed in compliance with the ESSCO Quality System manual, ECL1 Rev 46, dated 15 Feb 2022, ISO 9001:2015, ANSI/NCSL Z540-1-1994 Part 1, ISO 10012:2003, ISO 13485:2016, I SO/TS 16949:2009, and when required contractually, 10CFR21 and 10CFR50 App. B.

Amanda Matheson
Quality Manager