Essco Calibration Laboratory, LLC
27 Industrial Ave, Unit #9, Chelmsford, MA 01824-3618
Tel: 800-325-2201 978-250-0880 www.esscolab.com

Certificate of Calibration

Issue Date: 05/18/2026 **Certificate #: 3705471**

CUSTOMER / LOCATION

APPLI-TEC
7 INDUSTRIAL WAY
UNIT 1
SALEM, NH 03079

PURCHASE ORDER: 030220-00

EQUIPMENT INFORMATION

MANUFACTURER: JOHNSONITE
MODEL #: IQ TORO SC
SERIAL #: N/A
CONTROL #: 0792
TYPE: ESD FLOOR

PERFORMED: ONSITE

AS FOUND: IN TOLERANCE

AS LEFT: IN TOLERANCE

Remarks:

ENVIRONMENT: TEMPERATURE: 22.00 °C REL HUMIDITY: 48.00 %RH

CALIBRATION DATE: 05/15/2026 CALIBRATION DUE: 05/31/2027

Method			Description
ECP NO. 1.4.42	Rev: 4	12/01/2020	ESD Floors & Surfaces

Standard	Description	Report No.	Last Cal. Date	Cal. Due Date
E4709	DESCO 19291 METER, SURFACE RESISTANCE	3452136	6/24/2025	6/24/2026

The ESSCO Quality System is accredited to ISO/IEC 17025:2017

Ronny Noun
Metrologist

James Murphy
Quality Manager

The results above relate only to the item(s) calibrated. This certificate shall not be reproduced, except in full, without written approval of ESSCO. The signature to the right signifies responsibility for the quality system.

This document certifies that the unit conformed to applicable specifications upon successful completion of the calibration. Any number of factors may cause a nonconformance (i.e. out of tolerance) between calibrations. The measurement results are traceable to the International System of Units (SI Units) through a National Metrology Institute (NMI), such as NIST, a competent laboratory, or competent producer.

The calibration was performed in compliance with the ESSCO Quality System manual, ECL1 Rev 55, dated 16 April 2026, ISO 9001:2015, ANSI/NCSL Z540-1-1994 Part 1, ISO 10012:2026, ISO 13485:2016, and when required contractually, 10 CFR 21 and 10 CFR 50 App. B.